

# Certificate of Gauge Calibration

Issued by: **Wohler Retrotec EU B.V.**

Calibration Date: **2025-01-06**

Certificate Number: **402993 121644**

Results: **As Left**




Hardermaat 12

7244 PZ Barchem, The Netherlands

T: +31 522 282 941

E: calibration@retrotec.com W: www.retrotec.com



This calibration laboratory has been assessed by the ANSI National Accreditation Board and meets the requirements of international standard ISO/IEC 17025.

## Instrument:

Description: Pressure and Flow Gauge  
Manufacturer: Retrotec  
Model Number: DM32 10A  
Serial Number: 402993

Authorized by: Alex Peelle

Signature

## Environmental conditions:

Temperature: 20°C  $\pm$  6°C  
Relative Humidity: 50%  $\pm$  30%  
Mains Voltage: 120V  $\pm$  10V  
Mains Frequency: 60Hz  $\pm$  1Hz

Calibrated by: Max Kabel

Signature

Issue Date: 2025-01-06

## Comments:

Results recorded as received. No adjustment performed.

This calibration applies only to the unit listed on this certificate.

## Calibration Information:

The Device was calibrated against laboratory standards whose values are traceable to recognized national standards. The uncertainty represents an expanded uncertainty using a coverage factor of  $k=2$  to approximate a 95% confidence level. In tolerance conditions are based on test results falling within specified limits without taking uncertainty into account. The uncertainty evaluation has been carried out in accordance with ISO/IEC 17025 requirements.

## Calibration Procedure:

CP-35-01

This Calibration Certificate shall not be reproduced except in full, without written approval from Retrotec.

# Certificate of Gauge Calibration

Issued by: **Wohler Retrotec EU B.V.**  
Calibration Date: **2025-01-06**

Certificate Number: **402993 121644**  
Results: **As Left**

## Calibration Results

Channel A		
Applied Value (Pa)	Reading (Pa)	Error (%)
-2399.98	-2414.99	0.63
-1200.06	-1206.93	0.57
-599.99	-602.68	0.45
-300.00	-301.10	0.37
-100.00	-100.22	0.22
-75.00	-75.18	0.24
-50.00	-50.06	0.12
-24.99	-25.00	0.04
-9.95	-9.97	0.20
0.00	-0.01	N/A
10.01	10.10	0.90
24.99	25.11	0.48
50.01	50.24	0.46
75.02	75.32	0.40
100.01	100.37	0.36
300.00	300.85	0.28
600.05	601.13	0.18
1199.99	1200.90	0.08
2399.99	2400.21	0.01

Channel B		
Applied Value (Pa)	Reading (Pa)	Error (%)
-2399.98	-2420.68	0.86
-1200.06	-1208.83	0.73
-599.99	-603.47	0.58
-300.00	-301.42	0.47
-100.00	-100.39	0.39
-75.00	-75.19	0.25
-50.00	-50.10	0.20
-24.99	-25.01	0.08
-9.95	-9.98	0.30
0.00	0.00	N/A
10.01	10.09	0.80
24.99	25.15	0.64
50.01	50.21	0.40
75.02	75.37	0.47
100.01	100.46	0.45
300.00	301.04	0.35
600.05	601.54	0.25
1199.99	1201.75	0.15
2399.99	2401.37	0.06

Traceability is to the International System of Units (SI), consensus standards, or ratio type measurements through national standards realized and maintained by NIST or an NMI.

Instrument display resolution is 0.1 Pa.

## Uncertainties

Calibration and measurement capability (Expanded Uncertainty) is 0.073% of reading + 0.14 Pa (Range 0 - 2 000 Pa) based on a 95% confidence interval, using coverage of k=2.



Initial

# Certificate of Gauge Calibration

Issued by: **Wohler Retrotec EU B.V.**

Certificate Number: **402993 121644**

Calibration Date: **2025-01-06**

Results: **As Left**

The Pressure Range uncertainty limits are compliant with (meets or exceeds requirements from) the following standards or guidelines:

Name	Accuracy Requirements	Expiration Date
NFPA 2001	$\pm 1$ Pa $\pm$ (0 to 50 Pa)	2026-01-06
EN13829	$\pm 2$ Pa (up to $\pm 60$ Pa)	2030-01-06
ATTMA: TS1	$\pm 2$ Pa (up to $\pm 100$ Pa)	2026-01-06
ASTM E779-10	$\pm 5\%$ or 0.25 Pa whichever is greater	2030-01-06
CGSB	$\pm 1$ Pa (up to $\pm 60$ Pa)	2026-01-06
FD P50-784	$\pm 1\%$ or 1 Pa whichever is greater	2026-01-06
ISO14520	$\pm 1$ Pa (up to $\pm 60$ Pa)	2030-01-06
EN15004	$\pm 1$ Pa (up to $\pm 100$ Pa)	2030-01-06
USACE	$\pm 1\%$ or 0.25 Pa $\pm$ (25 to 250) Pa	2027-01-06
TITLE 24	$\pm 1\%$ or 0.2 Pa whichever is greater	2026-01-06
RESNET380	$\pm 1\%$ or 0.25 Pa whichever is greater	2026-01-06
ISO9972	$\pm 1$ Pa (up to $\pm 100$ Pa)	2030-01-06
FD E51-767	$\pm 2.5\%$ or 3 Pa whichever is greater	2027-01-06
RE2020	$\pm 3\%$ or 0.5 Pa $\pm$ (50 to 200) Pa	2027-01-06

End of report